

EC declaration of conformity

for the product **TRIO CARE**

We, Société Terraillon S.A.S., 1 rue Ernest Gouin, 78290 Croissy sur Seine, France, declare under our sole responsibility that product indicated above is compliant with essential requirements of health, safety and performances. This product complies with requirements of the **Directive 93/42/EEC concerning medical devices** and complies with the standards listed below:

Report TUV 50099483 001	IEC 60601-1-11 :2015 / EN60601-1-11 :2010
Report GZME170800071701	EN 60601-1-2 : 2014
Report QE-LT3062EPER02A01	IEC 60601-2-10 : 2012

Notified Body :

TÜV Rheinland LGA Product Gmbh
Tillystrasse 2 – 90431, Nürnberg, Germany
Certificat N°: HD60110846 0001
Issue date : 22 juin 2016
Expiry date 20 juin 2021

Manufacturer :

Shenzhen Dongdixin Technology Co, Ltd
N°. 3 Building XiliBaimang Xusheng
Industrial Estate, 518108 Nanshan
Shenzhen China

Authorized representative :

Shanghai international holding Corp. GmbH (Europe)
Eiffestrasse 80, 20537 Hamburg Germany

This product carries the CE mark **CE 0197**.

Croissy sur Seine, 02/07/2020

Cedric Dubois
Quality Director

