

## Declaration of Conformity

for the product

### Tensio Arm – PN 13829

We, Société Terraillon S.A.S., 1 rue Ernest Guoin, 78290 Croissy sur Seine, France, declare under our sole responsibility that product indicated above is compliant with essential requirements of health, safety and performances. This product complies with requirements of the **Directive 93/42/EEC concerning medical devices** and complies with the standards listed below:

Rapport GZME140600041702	IEC 60601-1-11 :2010
Rapport GMZE150600003301	EN 60601-1-2 : 2014
Rapport GMZE140600041704	EN 1060-3
Rapport GMZE140600041701	EN 60601-1-1
Rapport GMZE140600041704	EN 1060-1

**Notified Body :**

**TÜV SUD Product Service GmbH  
Ridlerstrasse 65 – 80339, Munchen, Germany**

This product carries the CE 0123 which was first affixed in 2019

Croissy sur Seine, 18th November 2020

Quality Manager  
Frederic DAGGAG

